

Technical Data Sheet

Doc No – Ab/TDS/CW-03

Rev No : 02

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Rack for Micro Centrifuge Tube, RPP

Valid for Catalog No. P20205

1. Description/Specification

1.1 Description:

Reusable Rack for Micro Centrifuge Tube, RPP

1.2 Design:

Compatible Rack for 1.5 ml and 2.0 ml Micro Centrifuge Tube

1.3 Performance Specifications:

P20205 (1.5ml & 2ml)

1.4 Material:

1.4.1 Comply with requirement of Directive EU No 10/2011 and requirement of FDA Regulation 21 CFR 177.1520 (c) 3.3b Polymers.

1.4.2 Materials are manufactured according to the requirements of GMP Regulation (EC) No 2023/2006

1.4.3 Material does not contain any Hazardous chemical substance & compliance with REACH AVHC & RoHS Free.

1.5 Color: White

1.6 Graduations:

Suitable for 1.5 & 2.0 ml Micro centrifuge Tube

1.7 Sterilization:

Can be autoclaved at 15 Psi, 121 °C for 20 minutes

1.8 Quality control:

Testing of various characteristic in accordance with the valid Specification

1.9 other Information:

Reusable, can be autoclaved upto to 10 autoclave cycle.



Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttarakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslaware.com
W : www.abdoslaware.com

Technical Data Sheet

1.9 Basic Features


1.9.1 Temperature range:

1.9.2 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

1.9.3 Chemical resistance: See website.

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.



Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttrakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslabware.com
W : www.abdoslabware.com