

Technical Data Sheet

Doc No – Ab/TDS/CT-010

Rev No : 04 Date 01/2021

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Rack for Micro Centrifuge Tube, PC

Valid for Catalog No. P20201, P20202, P20203, P20204

1. Description/Specification

1.1 Description:

Reusable Rack for Micro Centrifuge Tube, PC

1.2 Design:

Compatible Rack for 1.5 ml and 2.0 ml Micro Centrifuge Tube

1.3 Performance Specifications:

P20201 (1.5 & 2.0 ml), P20202 (1.5 & 2.0 ml), P20203 (1.5 & 2.0 ml), P20204 (1.5 & 2.0 ml)

1.4 Material:

1.4.2 Polycarbonate (PC), Confirm US FDA 21CFR

1.4.3 Complies with requirement accordance with article of regulation (EC) No. 1907/2006 and its amendments.

1.4.4 Complies with the requirement of European Pharmacopoeia, 7th edition (2011) & its supplements 7.5 (07/2012) monographs 3.2.2 Plastic Containers and closures for Pharmaceutical use.

1.4.5 The product has passed the USP testing including class VI tests & has been assigned the FDA Drug master file DMF 12123

1.4.6 Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.7 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE ,2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water

1.4.8 Product is manufactured without the use of Raw Materials of animal origin Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials



Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttarakhand, India
An ISO 9001 ,13485, 14001 & CE Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India
GST : 05AAGCA8605A1Z7

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com
CIN : U85100WB2008PTC123903

Technical Data Sheet

containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

1.5 Color: White & Yellow

1.6 Graduations:

Suitable for 1.5 & 2.0 ml Micro centrifuge Tube

1.7 Sterilization:

Can be autoclaved at 15 Psi, 121 °C for 20 minutes

1.8 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.9 other Information:

Reusable, can be autoclaved upto to 10 autoclave cycle.

2.0 Basic Features

2.1 Temperature range:

2.2 Autoclavability: Yes, Autoclavable at 15 psi and 121°C for 20 minutes

2.3 Manufactured in Aseptic environment

2.4 Chemical resistance: See website see website (www.abdoslifesciences.com)

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121°C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50°C before removing product.



Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttarakhand, India
An ISO 9001 ,13485, 14001 & CE Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India
GST : 05AAGCA8605A1Z7

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com
CIN : U85100WB2008PTC123903