

Technical Data Sheet

Doc No Ab/TDS/CT-19

Rev No: 02 Date 11/2019

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Swirl Pro High Speed Micro Centrifuge

Valid for Catalog No. E11520, E11530

1. Description Specifications:

1.1 Description: Swirl Pro High Speed Micro Centrifuge/ Swirl Refrigerated Pro High Speed Micro Centrifuge

1.1.1 Swirl Pro is a micro centrifuge used for sedimentation of cells and virus, separation of sub- cellular organelles such as DNA, RNA, proteins & lipids.

1.1.2 Swirl Pro has a large ergonomic design with easy to read display & sound alert.

1.1.3 Both RPM & G-force can be displayed.

1.1.4 Swirl Pro adapts a Brushless DC Motor.

1.1.5 Special multi flow air-cooling design helps to keep lower rotor temperature.

1.1.6 Special multi-flow air cooling design helps to keep lower rotor temp.

1.1.7 Unique cooling system design minimizes rotor heating to less than 10°C after running for one hour

1.2 Design: Swirl Pro High Speed Micro Centrifuge/ Swirl Refrigerated Pro High Speed Micro Centrifuge

1.3 performance Specifications:

Specifications	E11520	E11530
Voltage	Single-phase, AC 220-240V	Single-phase, AC 220-240V
Frequency	50-60 Hz	50-60 Hz
Current	5A	10A
Max Speed	15000 rpm (200-15000rpm) with increment of 10 rpm	15000 rpm (200-15000rpm) with increment of 10 rpm
Max RCF	21380xg with increment of 10xg	21380xg with increment of 10xg
Speed Accuracy	±20 rpm	±20 rpm



Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttrakhand, India
An ISO 9001 & 13485 Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com

Technical Data Sheet

Rotor Capacity	(i) 5ml x 18 places (ii) 0.2ml/0.5ml/1.5ml/2ml x 24 places (iii) 0.5ml x 36 places (iv) PCR 8x4 places	(i) 5ml x 18 places (ii) 0.2ml/0.5ml/1.5ml/2ml x 24 places (iii) 0.5ml x 36 places (iv) PCR 8x4 places
Run Time	30sec-99min- HOLD (Continuous operation)	30sec-99min- HOLD (Continuous operation)
Driving Motor	Brushless DC Motor	Brushless DC Motor
Safety Devices	(i) Dual door interlock (ii) Over- speed detection (iii) Over temperature detection (iv) Automatic internal diagnosis	(i) Dual door interlock (ii) Over- speed detection (iii) Over temperature detection (iv) Automatic internal diagnosis
Noise Level	≤ 58 dB	≤ 58 dB
Additional Features	(i) Speed/RCF switch (ii) Short-time run function (iii) Sound-alert function	(i) Speed/RCF switch (ii) Short-time run function (iii) Sound-alert function
Acceleration / Breaking time	25 sec for Both Acceleration & Breaking	25 sec for Both Acceleration & Breaking
Dimension	280x364x266 mm	388x580x324 mm
Weight	12 kg	30 kg
Permissible ambient Temp.	+4 to +40°C	+4 to +40°C
Permissible ambient Humidity	80%	80%

1.4 Material: ABS (Acrylonitrile Butadiene Styrene)

1.5 Color: White

1.6 E11520 – Swirl Pro High Speed Micro Centrifuge/ E11530-Swirl Refrigerated Pro High Speed Micro Centrifuge, CE certified

1.7 Spare Accessories

1.7.1 **E11521** – 24 Place Rotor for 0.2, 0.5, 1.5 & 2.0ml

1.7.2 **E11522** – 18 Place Rotor for 5.0ml

1.7.3 **E11523** - 36 Place Rotor for .02 & 5.0ml



Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttarakhand, India
An ISO 9001 & 13485 Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com

Technical Data Sheet

- 1.7.4 **E11524-** 32 (8x4) Place Rotor for 0.2 ml PCR
- 1.7.5 **E11525-** Rotor Adaptor for 2.0ml-0.2ml used with E11521
- 1.7.6 **E11526-** Rotor Adaptor for 2.0ml-0.5ml used with E11521
- 1.7.7 **E11527-** Rotor Adaptor for 0.5ml-0.2ml used with E11523
- 1.7.8 **E11528-** 12 Place Rotor for 5.0ml(conical bottom)



Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttrakhand, India
An ISO 9001 & 13485 Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com