Technical Data Sheet

Doc No- Ab/TDS/GLP-175

Rev No: 04 Date 02/2022-23

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Burette Clamp, PP

Valid for Catalog No. P70601, P70602

- 1. Description/Specification
- **1.1 Description:** Burette Clamp, PP
- 1.1.1 Burette clamps are used to hold burette, test tubes etc.
- 1.1.2 Completely corrosion proof as it is molded in polypropylene.
- 1.1.3 Spring loaded jack in the middle provides grip to the burette & prevents it from slipping.
- 1.1.4 Two tapered wedges are provided to ensure firm on the rods.
- 1.1.5 Available in both single & double clamp.

1.2 Design:

Burette Clamp, PP

1.3 Performance Specifications:

Code	Places	No. per Case
P70601	Single	6
P70602	Double	6

1.4 Material:

- 1.4.1 PP- Polypropylene (PP), confirm US FDA 21CFR
- 1.4.2 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.
- 1.4.3 The product has passed the USP testing including class VI tests, Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.
- 1.4.4 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations1935/2004/CE,2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water



Ph. No: +91- (0) 1332235277

Email:labtech@abdosindia.com

Abdos Labtech Private Limited

(A Divison of Abdos Group)

Technical Data Sheet

1.4.5 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

1.5 Color:

White

1.6 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release amendments.

- 1.7 Basic Features
- **1.7.1 Made up of** High pure virgin Polypropylene (PP)
- **1.7.2 Temperature range:** -80° C to $+121^{\circ}$ C
- 1.7.3 Autoclavability: Yes, Autoclavable at 15 psi and 121°C for 20 minutes
- **1.7.4 Chemical resistance:** See website (www.abdoslifesciences.com)
- 1.7.5 Manufactured in Aseptic Environment

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121^{0} C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50^{0} C before removing product.

Instruction for Use

Product can be used for food contact also.

Storage temp. Max 30°C

Before using the product check crackness etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.



Technical Data Sheet

WARNING

DO NOT store strong oxidizing agents.

DO NOT place any plastic labware in a flame.

DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure Consult local fire codes prior to storage of flammable liquids.







Ph. No: +91- (0) 1332235277 Email:labtech@abdosindia.com Web: www.abdoslifesiences.com GST: 05AAGCA8605A1Z7