

Technical Data Sheet

Doc No – Ab/TDS/CT-04
Rev No : 10 Date 04/2023-24

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Centrifuge Tubes Conical Bottom, PP/HDPE

Valid For Cat. No. P10401, P10402, P10403, P10404, P10405, P10406, P10407, P10408, P10430,
P10444, P10442, P10441, P10431, P10443, P10405T, P10406T, P10407T, P10408T, P10464

SRN (Single Registration Number): IN-MF-000022272

1. Description/Specification

1.1 Description:

Disposable PP/HDPE centrifuge tube, DNase, RNase, Human DNA, ATP & Pyrogen Free,
CE & IVD certified

1.2 Design:

1.2.1 Tube: Cylindrical Upper and conical Bottom

1.2.2 Tube Cap: Separate Threaded HDPE Cap

1.3 Performance Specifications:

Code	Capacity (ml)	OD (mm)	Height (mm)	Sterile	Colour	No./Case
P10430	15 Bulk	16.5	121	NO	Blue	500
P10401	15	16.5	121	NO	Blue	500
P10441	15	16.5	121	NO	Red	500
P10402	15	16.5	121	YES	Blue	500
P10442	15	16.5	121	YES	Red	500
P10431	50 Bulk	29.2	116	NO	Blue	500
P10403	50	29.2	116	NO	Blue	500
P10443	50	29.2	116	NO	Red	500
P10404	50	29.2	116	YES	Blue	500
P10444	50	29.2	116	YES	Red	500
P10405	15	16.5	121	NO	Tube-Amber, Cap - Blue	500
P10406	15	16.5	121	YES	Tube-Amber,	500



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakheswari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email: labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

					Cap - Blue	
P10407	50	29.2	116	NO	Tube-Amber, Cap - Blue	500
P10408	50	29.2	116	YES	Tube-Amber, Cap - Blue	500
P10405T	15	16.5	121	NO	TubeTr. Amber, Cap - Blue	500
P10406T	15	16.5	121	YES	Tube- Tr.Amber, Cap - Blue	500
P10407T	50	29.2	116	NO	Tube- Tr.Amber, Cap - Blue	500
P10408T	50	29.2	116	YES	Tube- Tr.Amber, Cap - Blue	500

1.4 Material:

1.4.1 Tube - Polypropylene (PP) & Cap- High Density Polyethylene (HDPE) confirm US FDA 21CFR

1.4.2 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.

1.4.3 The product has passed the USP testing including class VI tests. Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.4 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE, 2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water

1.4.5 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

1.5 Color:

Tube color- Translucent (Natural) Amber and Tr.Amber,

Cap color- Blue/Red

1.6 Graduation and writing surface:

1.6.1 Large Printed Graduation

1.6.2 White Patch for easy writing on the tube

1.6.3 Printed Max Freeze Level Indicator with Red Colour

1.7 Sterilization:

Sterile catalog number from this lot have been pre-sterilizes and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products Requirements for validation control- Radiation sterilization*. The product has been sterilized with Co-60 Gamma Radiation. SAL 10^{-6}



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakheswari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email: labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

1.8 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.9 Other Information

Recommended for single use only

2. Basic Features

2.1 Temperature range: Tube:-80⁰C to +121⁰C

2.2 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

2.3 Centrifugation: 18,000 RCF (15ml), 18,000 RCF (50ml conical)

2.4 Leak Test at 635 mm Hg: No leakage

2.5 Manufactured in Clean room ISO Class 8 (100,000)

2.6 Chemical resistance: see website (www.abdoslifesciences.com)

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.

*Depending on the quality of Centrifuge and compatibility of Rotor.

*Transparent Amber Centrifuge tubes can be supplied upon request.

Instruction for Use

Product can be used for food contact also.

Storage temp. Max 30⁰C

Before using the product check crackness, pouch punctured etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.

Ensure all sample tubes are evenly filled

For each tube inserted in the rotor, add a tube of equal weight directly opposite it.



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakheswari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email:labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

WARNING

DO NOT store strong oxidizing agents.


DO NOT place any plastic labware in a flame.

DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure

Wear a face shield and / or safety goggles if you have to work anywhere near a centrifuge that's in use.

Do not bump, jar, or move the centrifuge while the rotor is spinning. Make sure you don't have the cord dangling from a table edge where someone could catch their foot in it and pull down the centrifuge

Consult local fire codes prior to storage of flammable liquids.



Quality Manager

Abdos Labtech Private Limited

Employee Code 30110



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakshwari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903

Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email:labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7