Technical Data Sheet

Doc No Ab/TDS/ELECTRO-126

Rev No: 00 Date 02/2022-23

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

ABDOS Dual Wavelength UV Transilluminators

Valid for Catalog No. E12385, E12386, E12387

- 1. Description Specifications:
- **1.1 Description:** E12385 Dual wavelength UV Transilluminator, 254/365nm, 20x20cm
- E12386 Dual wavelength UV Transilluminator, 254/312nm, 20x20cm
- E12387 Dual wavelength UV Transilluminator, 254/302nm, 20x20cm
- 1.1.1 Abdos Dual wavelength UV transilluminators offers widest range of versatility with its three combinations of 254/365, 254/312 & 254/302nm.
 - 1.1.2 Compact 20x20cm size comes with UV protection filter cover for added safety.
- 1.1.3 Flicker free Eight UV tubes of 8 watts ensures even distribution of light for better viewing of EtBr Stained bands. Button to switch between the two wavelengths with markings.

1.2 Design: UV

1.3 Performance Specifications:

Specifications	E11385	E12386	E12387
Body	MS Powder Coated	MS Powder Coated	MS Powder Coated
Dimension (LxWxH)	35 x 30 x 15 cm	35 x 30 x 15 cm	35 x 30 x 15 cm
Viewing Area	20x20 cm	20x20 cm	20x20 cm
Wavelength	254nm	312nm	365nm
Detection Limit	1ng	1ng	1ng
No.of UV Tubes	8 Nos	8 Nos	8 Nos
Wattage	8 watt each	8 watt each	8 watt each
Filter	UV Protection	UV Protection	UV Protection

1.4 Material: Acrylic



GST: 05AAGCA8605A1Z7

Ph. No: +91- (0) 1332235277

Email:labtech@abdosindia.com Web: www.abdoslifesiences.com

Technical Data Sheet

1.5 CE Certified

1.6 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release

Warranty: 12 months from the date of Installation or 15 months from the date of Invoicing

Instruction for Use

Storage temp. Max 30°C

Before using the product check crackness etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.

WARNING

DO NOT store strong oxidizing agents.

DO NOT place any plastic labware in a flame.

DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure Consult local fire codes prior to storage of flammable liquids.





Registered office
13A Government

Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A. DDA

Technical Data Sheet





Ph. No: +91- (0) 1332235277 Email:labtech@abdosindia.com Web: www.abdoslifesiences.com GST: 05AAGCA8605A1Z7