

Technical Data Sheet

Doc No – Ab/TDS/MRB-125

Rev No : 05 Date 01/2021

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Rotator Work Station, PP

Valid for Catalog No. P20208

1. Description/Specification

1.1 Description: Reusable PP (Poly Propylene) Rotator Work Station

1.1.1 Abdos rotator workstation is a unique ultimate workstation helping you organize your laboratory work.

1.1.2 Cube shaped multi-hole rack suitable for different products, has a unique rotate and lock design that helps add or subtract racks according to end user's needs.

1.2 Design:

Cube Type Multi Hole for keeping Different Product as Per own requirement

1.3 Performance Specification:

1.3.1 Product	No. per unit
50 ml Centrifuge Tubes	2 per unit
15 ml Centrifuge Tube	8 per unit
1.5 & 2.5 Micro Centrifuge Tube	9 Per unit
0.5 ml Micro centrifuge tube	6 per unit
PCR Plates (Unskirted & semi Skirted 96 well)	1 per 2 units put together
PCR 8 Tubes strip	6 per unit
PCR 12 Tubes Strip	8 per 2 unit
PCR 0.2 ml Tubes	48 per unit

Universal fit all above Product as mentioned.

1.4 Material:

1.4.1 Polypropylene (PP), Hydrophobic, confirm US FDA 21CFR

1.4.2 Complies with requirement accordance with article of regulation (EC) No. 1907/2006 and its amendments.



Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttarakhand, India
An ISO 9001 ,13485, 14001 & CE Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India
GST : 05AAGCA8605A127

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com
CIN : U85100WB2008PTC123903

Technical Data Sheet

1.4.3 Complies with the requirement of European Pharmacopoeia, 7th edition (2011) & its supplements 7.5 (07/2012) monographs 3.2.2 Plastic Containers and closures for Pharmaceutical use.

1.4.4 The product has passed the USP testing including class VI tests & has been assigned the FDA Drug master file DMF 12123

1.4.5 Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.6 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE, 2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water

1.4.7 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

1.5 Color :

Assorted Colour - 4 Pcs per case – Blue, Green, Yellow, Red

1.6 Sterilization:

NA

1.7 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.8. Basic Features:

1.8.1 Made up of High pure virgin Polypropylene (PP)

1.8.2 Temperature range: -80⁰C to +121⁰C

1.8.3 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

1.8.4 Chemical resistance: see website (www.abdoslifesciences.com)

1.8.5 Interlocking Racks – Convenient to place all type of product in single racking products

1.8.6 Customize Specific requirements according to end user.

1.8.7 Rotate each of the four modules independently



Shiv Ganga Industrial Estate,

Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttarakhand, India

An ISO 9001 ,13485, 14001 & CE Certified

A Division of Abdos

Regd. Office:

13A, Government Place (East)
Kolkata – 700069, India

GST : 05AAGCA8605A127

P : +91 (0) 1332 235277

E : labtech@abdosindia.com

W: www.abdoslifesciences.com

CIN : U85100WB2008PTC123903

Technical Data Sheet


1.8.8 Rack maintains its orientation until twist again.

1.8.9 Dimension: 108 x 68 x108 mm

1.9 Manufactured in Aseptic Environment

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.


Quality Assurance Incharge
Abdos Labtech Pvt. Ltd.




ABDOS

Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttrakhand, India

An ISO 9001 ,13485, 14001 & CE Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India

GST : 05AAGCA8605A1Z7

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com
CIN : U85100WB2008PTC123903