

Technical Data Sheet

Doc No – Ab/TDS/Cryo-30

Rev No : 06 Date 04/2023-24

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Single Hand Cryo Holder Rack, PC

Valid for Catalog No. P20616

SRN (Single Registration Number): IN-MF-000022272

1. Description/Specification

1.1 Description:

Reusable Polycarbonate (PC) Single Hand Cryo Holder Rack.

1.2 Design:

Rectangular Design.

1.3 Performance Specification:

1.3.1 Capacity: P20616 – 1-2 ml Cryo tubes – 40 Places (10x4 Array).

1.3.2 Universal fit for 1-2 ml Cryo tubes Cryo tubes respectively.

1.3.3 Designed for use with cryogenic vial , freezer tubes storage vial and micro tubes.

1.3.4 The rack has a secure locking system in each well that will not allow the cryo vial to turn on either side.

1.3.5 Anti-skid rubber feet support the rack.

1.4 Material:

1.4.1 Polycarbonate (PC), confirm US FDA 21CFR

1.4.2 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.

1.4.3 The product has passed the USP testing including class VI tests ,Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.4 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE ,2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakshewari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email: labtech@abdosindia.com
Web: www.abdoslivesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

1.4.5 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

1.5 Color:

Rack: White, Yellow, Blue.

1.6 Sterilization:

Can be autoclaved at 15 Psi, 121 °C for 15 minutes

1.7 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.8 Other Information:

Reusable, can be autoclaved up to 10 autoclave cycle.

1.9. Basic Features

1.9.1 Made up of Polycarbonate (PC)

1.9.2 Performance: Boxes have hinged lid & ultra-smooth locking system ensure secure closure preventing accidental opening of these boxes in deep freezers.

1.9.3 Inbuilt numbered and letters grid for sample identification.

1.9.4 Temperature range: -196°C to +121°C

1.9.5 Autoclavability: Yes, Autoclavable at 15 psi and 121°C for 20 minutes

1.9.6 Manufactured in Aseptic Environment

1.9.7 Chemical resistance: see website (www.abdoslifesciences.com)

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121°C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50°C before removing product.

Instruction for Use

Product can be used for food contact also.

Storage temp. Max 30°C

Before using the product check crackness, pouch punctured etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakhewari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903

Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email:labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

WARNING

DO NOT store strong oxidizing agents.

DO NOT place any plastic labware in a flame.

DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure






Consult local fire codes prior to storage of flammable liquids.

UKRP
Obelis UK LTD
Sandford Gate,
East Point Business Park,
Oxford OX4 6LB
United Kingdom
Tel: +44 1491 378012
info@obelis.co.uk

Obelis s.a.
Bd General Wahis 53
1030 Brussels Belgium
Tel: +32 2 732-59-54
Fax: +32 2 732-60-03
mail@obelis.net

Abdos Labtech Pvt. Ltd.
Plot No. 1 Lakshwari Bhagwanpur Roorkee 247667 (UK) India
Email : labtech@abdosindia.com, www.abdoslifesciences.com
Ph: +91 11 41081897

Electronic If U



Quality Manager

Abdos Labtech Private Limited

Employee Code 30110

