

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-05
Rev No: 10 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Centrifuge Tubes Conical bottom, PS/HDPE

Valid For Cat. No. P10419, P10421

SRN (Single Registration Number): IN-MF-000022272

1. Description/Specification

1.1 Description:

Disposable PS (Polystyrene) centrifuge tube, DNase, RNase, Human DNA, ATP & Pyrogen Free, CE & IVD certified

1.2 Design:

1.2.1 Tube: Cylindrical Upper and conical Bottom

1.2.2 Tube Cap: Separate Threaded HDPE Cap

1.3 Performance Specifications:

Code	Capacity (ml)	OD (mm)	Height (mm)	Sterile	Colour	No./Case
P10419	15	16.5	121	YES	Blue	500 (Bulk)
P10421	15	16.5	121	YES	Blue	500(25pcsx20racks)

1.4 Material:

1.4.1 Tube - Polystyrene (PS) & Cap- High Density Polyethylene (HDPE) confirm US FDA 21CFR

1.4.2 The product complies with USP class VI. Free from Heavy metals, it is safe to use in contact with food stuff, Pharmaceuticals & drinking water.

1.4.3 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-05
Rev No: 10 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Centrifuge Tubes Conical bottom, PS/HDPE

1.5 Color:

Tube color- Natural

Cap color- Blue

1.6 Graduation and writing surface:

1.6.1 Large Printed Graduation

1.6.2 White Patch for easy writing on the tube

1.7 Sterilization:

Sterile catalog number from this lot have been pre-sterilizes and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products Requirements for validation and Routine control- Radiation sterilization*. The product has been sterilized with Co-60 Gamma Radiation. SAL 10^{-6}

1.8 Quality Control:

Products are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release

1.9 Other Information

Recommended for single use only

2. Basic Features

2.1 Temperature range: Tube:-10⁰C to +70⁰C

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-05
Rev No: 10 Date- 05/2024-25

Abdos Labtech Pvt. Ltd. Centrifuge Tubes Conical bottom, PS/HDPE

2.2 Autoclavability: NO, PS material can't be Autoclavable at doing such may cause deformation of shape.

2.3 Centrifugation: 2500 RCF (Rated with Double Distilled water at 25⁰C for 30 minutes)

2.4 Leak Test at 635 mm Hg: No leakage

2.5 These racks are made of eco-friendly paper to hold 25 pieces of CT 15 & 50ml.

2.6 Manufactured in Clean room ISO Class 8 (100,000)

2.7 Chemical resistance: see website (www.abdoslifesciences.com)

2.8 Intended use: Widely used in scientific studies of molecular biology, clinical chemistry, biochemistry and any type of scientific laboratory.

**Depending on the quality of Centrifuge and compatibility of Rotor.*

Autoclaving Guideline for Plastics

Polystyrene can't be autoclaved.

Instruction for Use

Product can be used for food contact also.

Storage temp. Max 30°C

Before using the product check crackness, pouch punctured etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.

Ensure all sample tubes are evenly filled

For each tube inserted in the rotor, add a tube of equal weight directly opposite it.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-05
Rev No: 10 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Centrifuge Tubes Conical bottom, PS/HDPE

WARNING

DO NOT store strong oxidizing agents.


DO NOT place any plastic labware in a flame.

DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure

Wear a face shield and / or safety goggles if you have to work anywhere near a centrifuge that's in use.

Do not bump, jar, or move the centrifuge while the rotor is spinning. Make sure you don't have the cord dangling from a table edge where someone could catch their foot in it and pull down the centrifuge

Consult local fire codes prior to storage of flammable liquids.



TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-05
Rev No: 10 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Centrifuge Tubes Conical bottom, PS/HDPE

Quality Manager
Abdos Labtech Private Limited
Employee Code 30110

