

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-06
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Centrifuge Tubes, Last Drop PP/HDPE

Valid For Cat. No. P10525, P10526, P10527, P10528

SRN (Single Registration Number): IN-MF-000022272

1. Description/Specification

1.1 Description:

Disposable PP/HDPE centrifuge tube, DNase, RNase, Human DNA, ATP & Pyrogen Free, CE & IVD certified

1.2 Design:

1.2.1 Tube: Cylindrical Upper and conical Bottom

1.2.2 Tube Cap: Separate Threaded HDPE Cap

1.3 Performance Specifications:

Code	Capacity (ml)	OD (mm)	Height (mm)	Sterile	Colour	No./Case
P10525	15	16.5	121	NO	Natural	500
P10526	15	16.5	121	YES	Natural	500
P10527	50	29.2	116	NO	Natural	500
P10528	50	29.2	116	YES	Natural	500

1.4 Material:

1.4.1 Tube - Polypropylene (PP) & Cap- High Density Polyethylene (HDPE) confirm US FDA 21CFR

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-06
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd. Centrifuge Tubes, Last Drop PP/HDPE

1.4.2 The product complies with USP class VI. Free from Heavy metals, it is safe to use in contact with food stuff, Pharmaceuticals & drinking water.

1.4.3 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

1.5 Color:

Tube color- Natural

Cap color- Natural

1.6 Graduation and writing surface:

1.6.1 Large Printed Graduation

1.6.2 White Patch for easy writing on the tube

1.6.3 Printed Max Freeze Level Indicator with Red Colour

1.6.4 Abdos low binding technology provides 10 times lesser retention level with highly viscous samples & saves costly reagents. Super homogenous surface free sites prevent non-specific binding.

1.6.5 Reverse Graduation

1.7 Sterilization:

Sterile catalog number from this lot have been pre-sterilizes and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control- Radiation sterilization*. The product has been sterilized with Co-60 Gamma Radiation. SAL 10^{-6}

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-06
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd. Centrifuge Tubes, Last Drop PP/HDPE

1.8 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.9 Other Information

Recommended for single use only

2. Basic Features

2.1 Temperature range: Tube:-80⁰C to +121⁰C

2.2 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

2.3 Centrifugation: 18000 RCF (Rated with Double Distilled water at 10⁰C for 30 minutes)

2.4 Leak Test at 635 mm Hg: No leakage

2.5 Manufactured in Clean room ISO Class 8 (100,000)

2.6 Chemical resistance: see website (www.abdoslifesciences.com)

2.7 Intended use: Widely used in scientific studies of molecular biology, clinical chemistry, biochemistry and any type of scientific laboratory.

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-06
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd. Centrifuge Tubes, Last Drop PP/HDPE

Instruction for Use

Product can be used for food contact also.
Storage temp. Max 30°C
Before using the product check crackness, pouch punctured etc.
Check all the information printed on pouches or boxes.
Always wear gloves while handling specimen.
Ensure all sample tubes are evenly filled
For each tube inserted in the rotor, add a tube of equal weight directly opposite it.


WARNING

DO NOT store strong oxidizing agents.
DO NOT place any plastic labware in a flame.
DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure
Wear a face shield and / or safety goggles if you have to work anywhere near a centrifuge that's in use.
Do not bump, jar, or move the centrifuge while the rotor is spinning. Make sure you don't have the cord dangling from a table edge where someone could catch their foot in it and pull down the centrifuge
Consult local fire codes prior to storage of flammable liquids.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-06
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Centrifuge Tubes, Last Drop PP/HDPE



Quality Manager

Abdos Labtech Private Limited

Employee Code 30110

