

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-10
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Screw Cap Tube Rack, PC

Valid for Catalog No. P20235

SRN (Single Registration Number): IN-MF-000022272

1. Description/Specification

1.1 Description:

Reusable Rack for Screw Cap Tube, PC

1.2 Design:

Compatible Rack for 0.5/1.5- and 2.0-ml Screw Cap Tube

1.3 Performance Specifications:

Code	Capacity(ml)	LxWxH(mm)	Place	No./Case
P20235	0.5/1.5/2.0	256x61x41	20 (10x2)	1

1.4 Material:

1.4.1 Polycarbonate (PC), Confirm US FDA 21CFR

1.4.2 The product complies with USP class VI. Free from Heavy metals, it is safe to use in contact with food stuff, Pharmaceuticals & drinking water.

1.4.3 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-10
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Screw Cap Tube Rack, PC

1.5 Color: Transparent

1.6 Graduations:

Indexing 1 to 10

1.7 Sterilization:

NA

1.8 Quality Control:

Products are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.9 Other Information:

Reusable, can be autoclaved upto to 10 autoclave cycle.

2.0 Basic Features

2.1 Temperature range: Tube:-40⁰C to +100⁰C

2.2 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

2.3 Manufactured in Aseptic environment.

2.4 Chemical resistance: See website see website (www.abdoslifesciences.com)

2.5 Knurl Locking for one hand operations.

2.6 Intended use: Widely used in scientific studies of molecular biology, clinical chemistry, biochemistry and any type of scientific laboratory.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-10
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Screw Cap Tube Rack, PC

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.

Instruction for Use

Product can be used for food contact also.
Storage temp. Max 30°C
Before using the product check crackness, pouch punctured etc.
Check all the information printed on pouches or boxes.
Always wear gloves while handling specimens.






WARNING

DO NOT store strong oxidizing agents.
DO NOT place any plastic labware in a flame.
DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure
Consult local fire codes prior to storage of flammable liquids.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/CT-10
Rev No: 02 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Screw Cap Tube Rack, PC

   		UKRP Obelis UK LTD Sandford Gate, East Point Business Park, Oxford OX4 6LB United Kingdom Tel: +44 1491 378012 info@obelis.co.uk UK Obelis s.a. Bd General Wabis 53 1030 Brussels Belgium Tel: +(32) 2 732-59-54 Fax: +(32) 2 732-60-03 mail@obelis.net
Manufactured & Marketed By : Abdos Labtech Private Limited Plot No. 1 Lakshewari Bhagwanpur Roorkee 247667 (UK) India For Complaints/Feedback Contact Customer Care at above address or Ph: +91-011-41081897 or Email us at labtech@abdosindia.com Lic. No.: MFG/IVD/2024/000115 MADE IN INDIA		 Electronic FU Scan to visit website

Quality Manager

Abdos Labtech Private Limited

Employee Code 30110

