

TECHNICAL DATA SHEET

Doc No – Ab/TDS/LH-41
Rev No: 07 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Pipette Rack Stand (Carosol Type), PP

Valid for Catalog No. P70203

SRN (Single Registration Number): IN-MF-000022272

1. Description/Specification

1.1 Description: Pipettes Rack Stand, PP

1.1.1 Keep your pipettes within easy inside this convenient rack.

1.1.2 Holds six standard size pipettes.

1.1.3 Racks are mounted on their own stand.

1.1.4 Keep working area or desk free from clutter.

1.2 Design: Pipette Rack Stand (Carosol Type)

1.3 Performance Specifications:

Code	Places	No. per Case
P70203	6	1

1.4 Material:

1.4.1 Polypropylene (PP), Hydrophobic, confirm US FDA 21CFR

1.4.2 Free from Heavy metals, It is safe use in contact with food stuff, Pharmaceuticals & drinking water.

1.4.3 Product is manufactured without the use of Raw Materials of animal origin Therefore BSE/TSE issue does not concern this product to the best of our knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/LH-41
Rev No: 07 Date- 05/2024-25

Abdos Labtech Pvt. Ltd. Pipette Rack Stand (Carosol Type), PP

1.5 Color: Transparent

1.6 Quality control: Products are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.7. Basic Features

1.7.1 Made up of Polypropylene (PP)

1.7.2 Strong & Lightweight material.

1.7.3 Good impact strength, higher than both glass and polystyrene.

1.7.4 Fitment with ABDOS Pipettes only

1.7.5 **Autoclavability:** Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

1.7.6 Manufactured in Aseptic Environment

1.7.7 **Chemical resistance:** see website (www.abdoslifesciences.com)

1.8 Intended use: Widely used in scientific studies of molecular biology, clinical chemistry, biochemistry and any type of scientific laboratory.

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.

TECHNICAL DATA SHEET

Doc No – Ab/TDS/LH-41
Rev No: 07 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Pipette Rack Stand (Carosol Type), PP

Instruction for Use

Product can be used for food contact also.
Storage temp. Max 30°C
Before using the product check crackness etc.
Check all the information printed on pouches or boxes.
Always wear gloves while handling specimens.

WARNING

DO NOT store strong oxidizing agents.
DO NOT place any plastic labware in a flame.
DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure.
Consult local fire codes prior to storage of flammable liquids.



Quality Manager
Abdos Labtech Private Limited
Employee Code 30110



TECHNICAL DATA SHEET

Doc No – Ab/TDS/LH-41
Rev No: 07 Date- 05/2024-25

Abdos Labtech Pvt. Ltd.
Pipette Rack Stand (Carosol Type), PP

