

# Technical Data Sheet

Doc No – Ab/TDS/Cryo-25

Rev No : 08 Date 02/2022-23

## TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Freezing Tubes , PP

Valid for Cat. No. P60102, P60103, P60102NP, P60102NPNS

### 1. Description/Specification

**1.1 Description:** Freezing Cryo vials, DNase, RNase, Human DNA, ATP & Pyrogen Free, IVD certified

#### 1.2 Design:

1.2.1 Vials: Cylindrical External Threads with skirted foot (PP)

1.2.3 Cap: Assembled (HDPE)

#### 1.3 Performance Specifications:

1.3.1 Vial Volume: 1.8 ml and  $4.5 \pm 0.1$  ml

1.3.2 Vial Size : (OD) 11.80 x (H) 45.7 mm (P60102), 11.80 x 86.5 mm (P60103),

Cap Size : 13.2 mm ( $\phi$ )

P60102NP – without print Sterile (2pouchx500pcs) = 1000pcs/case

P60102NPNS – without print Non-Sterile (2pouchx500pcs) = 1000pcs/case

Code	Capacity (ml)	Outer Diameter (mm)	Height	No. Per Case	Self-Standing	Bottom	Brimful Capacity (Approx ml)
P60102	1.8	11.80	45.7	1000	Yes	Skirted	2.80
P60103	4.5	11.80	86.5	500	Yes	Skirted	5.30

#### 1.3.3 Graduation and writing surface:

1.3.3.1 Large Printed Graduation

1.3.3.2 White Patch for easy writing on the tube

#### 1.4 Material:

1.4.1 **Tube:** Polypropylene (PP), Hydrophobic **Cap**-HDPE

1.4.2 Polypropylene (PP), Hydrophobic, confirm US FDA 21CFR



Abdos Labtech Private Limited  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakheswari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

Registered office  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903  
Corporate Office  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277  
Email: labtech@abdosindia.com  
Web: [www.abdoslivesciences.com](http://www.abdoslivesciences.com)  
GST: 05AAGCA8605A1Z7

# Technical Data Sheet

1.4.3 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.

1.4.4 The product has passed the USP testing including class VI tests, Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.5 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE, 2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water

1.4.6 Product is manufactured without the use of Raw Materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

## 1.5 Color:

Translucent (Natural)

## 1.6 Sterilization:

Sterile catalog number from this lot have been pre-sterilizes and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control- Radiation sterilization*. The product has been sterilized with Co-60 Gamma Radiation. SAL  $10^{-6}$

## 1.7 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

## 1.8 Other information:

Recommended for single use only

**Vial Pack Size:** 1000 pieces/case (P60102) & 500 pieces/case (P60103)

## 2. Basic Features

**2.1 Temperature range:** Tube:  $-80^{\circ}\text{C}$  to  $+121^{\circ}\text{C}$  (Can withstand in liquid nitrogen with standard procedure, recommended to store in vapor phase of liquid nitrogen).

**2.2 Autoclavability:** Autoclavable at 15 psi and  $121^{\circ}\text{C}$  for 20 minutes

**2.3 Leak Test at 635 mm Hg with Cap fitted:** No leakage

**2.4** Manufactured in Clean room ISO Class 8 (100,000)

**2.5 Chemical resistance:** see website ([www.abdoslifesciences.com](http://www.abdoslifesciences.com))



Abdos Labtech Private Limited  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakshewari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

**Registered office**  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903  
**Corporate Office**  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277  
Email: [labtech@abdosindia.com](mailto:labtech@abdosindia.com)  
Web: [www.abdoslifesciences.com](http://www.abdoslifesciences.com)  
GST: 05AAGCA8605A1Z7

# Technical Data Sheet

## Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121<sup>0</sup>C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50<sup>0</sup>C before removing product.

## Instruction for Use

Product can be used for food contact also.

Storage temp. Max 30<sup>0</sup>C

Before using the product check crackness, pouch punctured etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.

## WARNING

**DO NOT** store strong oxidizing agents.

**DO NOT** place any plastic labware in a flame.

**DO NOT** mix any chemicals that may result in a thermic reaction, which can cause product failure

Consult local fire codes prior to storage of flammable liquids.

UKRP  
Obelis UK LTD  
Sandford Gate,  
East Point Business Park,  
Oxford OX4 5LB  
United Kingdom  
Tel: +44 1491 378012  
info@obelis.co.uk

UK CA

30<sup>0</sup>C

STERILE R

CE

IVD

EC REP

Obelis s.a.  
80 General Walleis 53  
1030 Brussels Belgium  
Tel: +(32) 2 732-59-54  
Fax: +(32) 2 732-60-03  
mail@obelis.net

**Abdos Labtech Pvt. Ltd.**  
Plot No. 1 Lakheshwari Bhagwanpur Roorkee 247667 (UK) India  
Email : labtech@abdosindia.com, www.abdoslifesciences.com  
Ph: +[91] 11 41081897

Electronic EU

Quality Assurance Incharge  
Abdos Labtech Pvt. Ltd.



**Abdos Labtech Private Limited**  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakheshwari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

**Registered office**  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903  
**Corporate Office**  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277  
Email:labtech@abdosindia.com  
Web: [www.abdoslifesciences.com](http://www.abdoslifesciences.com)  
GST: 05AAGCA8605A1Z7

# Technical Data Sheet



P60102

P60103



**Abdos Labtech Private Limited**  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakheswari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

**Registered office**  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903

**Corporate Office**  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

**Ph. No: +91- (0) 1332235277**  
**Email:labtech@abdosindia.com**  
**Web: [www.abdoslivesciences.com](http://www.abdoslivesciences.com)**  
**GST: 05AAGCA8605A1Z7**