

Technical Data Sheet

Doc No – Ab/TDS/CT-02

Rev No: 05 Date 04/2023-24

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Last Drop Low Retention Micro Centrifuge Tube, PP

Valid for Cat. No. P10208, P10209, P10210

SRN (Single Registration Number): IN-MF-000022272

1. Description/Specification

1.1 Description:

Disposable PP Micro Centrifuge tube, DNase, RNase, Human DNA, ATP & Pyrogen Free , CE & IVD certified

1.2 Design:

1.2.1 Tube: Cylindrical upper and conical bottom

1.2.2 Tube cap: Attached Snap Cap with lid open force kgf

1.3 Performance Specification:

1.3.1 Super hydrophobic material confirming retention level 10 times lesser & resulting in a better sample recovery.

1.3.2 These tubes come with integral snap cap.

Code	Capacity (ml)	OD (mm)	Height (mm)	Sterile	No./Case
P10208	0.5	10	31.5	NO	1000
P10208	1.5	13	41	NO	500
P10208	2.0	13	41	NO	500

1.4 Material:

1.4.1 Polypropylene (PP), Hydrophobic, confirm US FDA 21CFR

1.4.2 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.

1.4.3 The product has passed the USP testing including class VI tests. Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.4 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE ,2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakheswari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email: labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

1.4.5 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

1.5 Color:

Translucent (Natural)

1.6 Graduation and Writing Surface

1.6.1 Large Molded Graduation

1.6.2 Frosted Surface for easy writing on tube

1.7 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.8 Other Information

DNase, RNase, Human DNA, ATP & Pyrogen Free

Recommended for Single use only

1.9 Basic Features

1.9.1 Temperature range: Tube:-80 °C to + 121°C

1.9.2 Autoclavability: Yes, Autoclavable at 15 psi and 121 °C for 20 minutes

1.9.3 Centrifugation: 14,000* to 20,000 RCF (* Rated with Phenol/ Chloroform)

1.9.4 Leak Test at 635 mm Hg: No leakage

1.9.5 Boiling test : Pass at 100°C for 30 minutes in water bath.

1.9.6 Manufactured in Clean room ISO Class 8 (100,000)

1.9.6 Chemical resistance: see website (www.abdoslifesciences.com)

Autoclaving Guideline for Plastics

All plastic consumable are Autoclavable at 121°C for 20 minutes, 15 psi/1atm. Do not use dry cycle as this may cause deformation or weakening of plastic. Allow temperature within the autoclave to return to at least 50°C before removing product.

Instruction for Use

Product can be used for food contact also.

Storage temp. Max 30°C

Before using the product check crackness, pouch punctured etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.

Ensure all sample tubes are evenly filled

For each tube inserted in the rotor, add a tube of equal weight directly opposite it.



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakheswari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email:labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

WARNING

DO NOT store strong oxidizing agents.


DO NOT place any plastic labware in a flame.

DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure

Wear a face shield and / or safety goggles if you have to work anywhere near a centrifuge that's in use.

Do not bump, jar, or move the centrifuge while the rotor is spinning. Make sure you don't have the cord dangling from a table edge where someone could catch their foot in it and pull down the centrifuge

Consult local fire codes prior to storage of flammable liquids.



Quality Manager

Abdos Labtech Private Limited

Employee Code 30110



PIQ208

PIQ209



PIQ210



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakhewari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903

Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email:labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7