

# Technical Data Sheet

Doc No Ab/TDS/CT-18

Rev No: 04 Date 10/2021-22

## TECHNICAL DATA SHEET

**Abdos Labtech Pvt. Ltd.**

**Swirl Pro High Speed Micro Centrifuge**

**Valid for Catalog No. E11520**

### **1. Description Specifications:**

#### **1.1 Description: Swirl Pro High Speed Micro Centrifuge**

1.1.1 Swirl Pro is a micro centrifuge used for sedimentation of cells and virus, separation of sub-cellular organelles such as DNA, RNA, proteins & lipids.

1.1.2 Swirl Pro has a large ergonomic design with easy to read display & sound alert.

1.1.3 Both RPM & G-force can be displayed.

1.1.4 Swirl Pro adapts a Brushless DC Motor.

1.1.5 Special multi flow air-cooling design helps to keep lower rotor temperature.

1.1.6 Special multi-flow air cooling design helps to keep lower rotor temp.

1.1.7 Unique cooling system design minimizes rotor heating to less than 10°C after running for one hour

#### **1.2 Design: Swirl Pro High Speed Micro Centrifuge**

#### **1.3 performance Specifications:**

<b>Specifications</b>	<b>E11520</b>
<b>Voltage</b>	Single-phase, AC 220-240V
<b>Frequency</b>	50-60 Hz
<b>Current</b>	5A
<b>Max Speed</b>	15000 rpm (200-15000rpm) with increment of 10 rpm
<b>Max RCF</b>	21380xg with increment of 10xg
<b>Speed Accuracy</b>	±20 rpm
<b>Rotor Capacity</b>	(i) 5ml x 18 places



**Shiv Ganga Industrial Estate,**  
Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttarakhand, India  
**An ISO 9001 ,13485, 14001 & CE Certified**  
A Division of Abdos

**Regd. Office:**  
13A, Government Place (East)  
Kolkata – 700069, India  
**GST : 05AAGCA8605A1Z7**

**P : +91 (0) 1332 235277**  
**E : labtech@abdosindia.com**  
**W: www.abdoslifesciences.com**  
**CIN : U85100WB2008PTC123903**

# Technical Data Sheet

	(ii) 0.2ml/0.5ml/1.5ml/2ml x 24 places (iii) 0.5ml x 36 places (iv) PCR 8x4 places
<b>Run Time</b>	30sec-99min- HOLD (Continuous operation)
<b>Driving Motor</b>	Brushless DC Motor
<b>Safety Devices</b>	(i) Dual door interlock (ii) Over- speed detection (iii) Over temperature detection (iv) Automatic internal diagnosis
<b>Noise Level</b>	≤ 58 dB
<b>Additional Features</b>	(i) Speed/RCF switch (ii) Short-time run function (iii) Sound-alert function
<b>Acceleration / Breaking time</b>	25 sec for Both Acceleration & Breaking
<b>Dimension</b>	280x364x266 mm
<b>Weight</b>	12 kg
<b>Permissible ambient Temp.</b>	+4 to +40°C
<b>Permissible ambient Humidity</b>	80%

**1.4 Material:** ABS (Acrylonitrile Butadiene Styrene)

**1.5 Color:** White

**1.6 E11520** – Swirl Pro High Speed Micro Centrifuge, CE certified

**1.7 Spare Accessories**

1.7.1 **E11521** – 24 Place Rotor for 0.2, 0.5, 1.5 & 2.0ml

1.7.2 **E11522** – 18 Place Rotor for 5.0ml

1.7.3 **E11523** - 36 Place Rotor for .02 & 5.0ml

1.7.4 **E11524**- 32 (8x4) Place Rotor for 0.2 ml PCR



**Shiv Ganga Industrial Estate,**  
Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttarakhand, India

**An ISO 9001 ,13485, 14001 & CE Certified**

A Division of Abdos

**Regd. Office:**  
13A, Government Place (East)  
Kolkata – 700069, India

**GST : 05AAGCA8605A1Z7**

**P** : +91 (0) 1332 235277

**E** : labtech@abdosindia.com

**W**: www.abdoslifesciences.com

**CIN** : U85100WB2008PTC123903

# Technical Data Sheet

1.7.5 **E11525**- Rotor Adaptor for 2.0ml-0.2ml used with E11521

1.7.6 **E11526**- Rotor Adaptor for 2.0ml-0.5ml used with E11521

1.7.7 **E11527**- Rotor Adaptor for 0.5ml-0.2ml used with E11523

**Warranty:** 12 months from the date of Installation or 15 months from the date of Invoicing.



**Shiv Ganga Industrial Estate,**  
Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttrakhand, India

**An ISO 9001 ,13485, 14001 & CE Certified**

A Division of Abdos

**Regd. Office:**

13A, Government Place (East)  
Kolkata – 700069, India

**GST : 05AAGCA8605A1Z7**

**P : +91 (0) 1332 235277**

**E : labtech@abdosindia.com**

**W: www.abdoslifesciences.com**

**CIN : U85100WB2008PTC123903**