

# Technical Data Sheet

Doc No – Ab/TDS/MRB-129

Rev No: 03 Date 01/2021

## TECHNICAL DATA SHEET

**Abdos Labtech Pvt. Ltd.**

**Twisty Tube Rack, PP**

**Valid for Cat. No. P20210**

### **1. Description/Specification:**

#### **1.1 Description:** Twisty Tube Rack, PP

1.1.1 These racks make dilution and concentration study easier as it can transfer samples between 0.2ml PCR tubes to 0.5, 1.5 & 2.0ml tubes.

1.1.2 Can be rotated and customize according to end user application.

1.1.3 These racks have tapered wells that securely suspend tubes and makes removal easy.

1.1.4 Molded in indexing for easy retrieval of tubes.

#### **1.2 Design:** Twisty Tube Rack

#### **1.3 Performance Specifications:**

<b>Code</b>	<b>L x W x H (mm)</b>	<b>No. Per Case</b>
<b>P20210</b>	296x118x34	1 Kit (Blue + Purple)

<b>Color</b>	<b>Product</b>	<b>Places</b>
<b>Blue</b>	0.5 ml	48
	1.5/2.0ml	48
<b>Purple</b>	0.2ml	96
	1.5/2.0ml	48

#### **1.4 Material:**

1.4.1 Polypropylene (PP), Hydrophobic, confirm US FDA 21CFR

1.4.2 Complies with requirement accordance with article of regulation (EC) No. 1907/2006 and its amendments.



**Shiv Ganga Industrial Estate,**  
Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttarakhand, India  
**An ISO 9001 ,13485, 14001 & CE Certified**  
A Division of Abdos

**Regd. Office:**  
13A, Government Place (East)  
Kolkata – 700069, India  
**GST : 05AAGCA8605A127**

**P : +91 (0) 1332 235277**  
**E : labtech@abdosindia.com**  
**W: www.abdoslifesciences.com**  
**CIN : U85100WB2008PTC123903**

# Technical Data Sheet

1.4.3 Complies with the requirement of European Pharmacopoeia, 7<sup>th</sup> edition (2011) & its supplements 7.5 (07/2012) monographs 3.2.2 Plastic Containers and closures for Pharmaceutical use

1.4.4 The product has passed the USP testing including class VI tests & has been assigned the FDA Drug master file DMF 12123

1.4.5 Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.6 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE, 2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water

1.4.7 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product

**1.5 Color:** Blue & Purple

**1.6 Sterilization:**

NA

**1.7 Quality control:**

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

**1.8 Other information**

Reusable, can be used in wide temperature range

**1.9 Basic Features**

**1.9.1 Temperature range:** -80<sup>0</sup>C to +121<sup>0</sup>C

**1.9.2 Autoclavability:** Yes, Autoclavable at 15 psi and 121<sup>0</sup>C for 20 minutes

**1.9.3 Chemical resistance:** See website ([www.abdoslifesciences.com](http://www.abdoslifesciences.com))

**1.9.4** Manufactured in Aseptic Environment



**Shiv Ganga Industrial Estate,**

Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttarakhand, India

**An ISO 9001 ,13485, 14001 & CE Certified**

A Division of Abdos

**Regd. Office:**

13A, Government Place (East)  
Kolkata – 700069, India

**GST : 05AAGCA8605A1Z7**

**P : +91 (0) 1332 235277**

**E : labtech@abdosindia.com**


**W: www.abdoslifesciences.com**

**CIN : U85100WB2008PTC123903**

# Technical Data Sheet

## Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121<sup>0</sup>C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50<sup>0</sup>C before removing product.



**Shiv Ganga Industrial Estate,**  
Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttrakhand, India  
**An ISO 9001 ,13485, 14001 & CE Certified**  
A Division of Abdos

**Regd. Office:**  
13A, Government Place (East)  
Kolkata – 700069, India  
**GST : 05AAGCA8605A1Z7**

**P : +91 (0) 1332 235277**  
**E : labtech@abdosindia.com**  
**W: www.abdoslifesciences.com**  
**CIN : U85100WB2008PTC123903**